

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130307
Name		JAMES I. JENSEN, JR	ID# 3393 Date 6-2-09
A	Agency	YAKUTAT DPS	Phone # 784-3206
Instrument Location		AT YDPS	✓
Alco S/N		X124635	Target Value .079 High Pressure 810
Alco Test Values		.081	.080
		1 st Alco	2 nd Alco
Signature		[Signature]	
		(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 12 day of June 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130307

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

JUNE 03, 2009 ✓ 06/03/09

OPERATOR'S NAME:

JENSEN JAMES I

OPERATOR'S NUMBER: 3393

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: YAK1

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X124635 ✓

— BREATH ANALYSIS —

.079 ADJUSTED FOR 30.51 in ✓
ALCO TARGET .000 00:45
BLANK TEST .000 00:46
INTERNAL STANDARD VERIFIED 00:46
ALCO TV 30.51 in .001 ✓ 00:47
BLANK TEST .000 00:48
SUBJECT SAMPLE 06/03/09 .000 00:49
BLANK TEST .000 00:50
ALCO TV 30.51 in .000 ✓ 00:50
BLANK TEST .000 00:51

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

JUNE 03, 2009 ✓
TIME 00:43

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/28/01

HEATERS

SAMPLE CHAMBER: 51c

BREATH TUBE: 42c

BAROMETER: 30.51 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@AB CDEFG
HIJKLMN O PQRSTU VWXYZ[\]^_`abcdef ghijklmno
pqrstuvwxyz{|}~■